

Weighted Blanket
 Serial #XL309-U880NR7407 Site: Date: 11/15/2007

No	Site	Insp	Flr	Side	Room	Source	Sub	Feat	Cnd	Clr	Ssec	Date/Time	DI	Result	Pbl ± Prec	Pbk ± Prec	Pbc ± Prec
1					Shutter Cal 1						67.9	11/15/2007 13:38:54	0.0	...	NA	NA	NA
2		KJM			Calibrate 0.0						3.2	11/15/2007 13:40:55	1.0	NEG	0.01 ± 0.21	-1.59 ± 1.42	0.01 ± 0.21
3		KJM			Calibrate 0.3						3.2	11/15/2007 13:41:07	1.0	NEG	0.28 ± 0.17	0.54 ± 1.59	0.28 ± 0.17
4		KJM			Calibrate 1.0						16.4	11/15/2007 13:41:23	1.1	INCOM	1.13 ± 0.16	0.87 ± 0.61	1.13 ± 0.16
5		KJM			Calibrate 1.6						11.9	11/15/2007 13:42:12	1.2	POS	1.46 ± 0.24	0.65 ± 0.81	1.46 ± 0.24
6		KJM			Blanket	Blanket		Blanket	Intact	Pink	5.4	11/15/2007 13:52:15	2.2	POS	>>5.0	44.65 ± 4.61	44.65 ± 4.61
7		KJM			Blanket	Blanket		Blanket	Intact	Pink	5.0	11/15/2007 13:55:52	1.7	POS	>>5.0	66.76 ± 6.21	66.76 ± 6.21