

ENVIRONMENTAL HAZARDS SERVICES, L.L.C.

7469 WHITE PINE ROAD - RICHMOND, VA 23237

804-275-4788 FAX 804-275-4907

WIPE LEAD ANALYSIS SUMMARY

CLIENT: St. Clair County Health Dept.
3415 28th Street
Port Huron, MI 48060

DATE OF SAMPLING: 15 Nov 2007
DATE OF RECEIPT: 19 Nov 2007
DATE OF ANALYSIS: 19 Nov 2007
DATE OF REPORT: 20 Nov 2007

CLIENT NUMBER: 23-6245 D
EHS PROJECT #: 2007-11-1767
PROJECT: 0001

EHS SAMPLE#	CLIENT SAMPLE#	WIPE AREA (ft ²)	TOTAL LEAD (ug)	CONCENTRATION (ug/ft ²)
01	D1	1.00	160	160
02	D2	-----	<20	-----
03	D3	0.250	67	270


QUALITY CONTROL DATA

BATCH#:	111907W-20
INCLUSIVE EHS SAMPLE NUMBERS:	01-03
Continuing Calibration Verification 10 (10.0ppm Pb)	98.2% Recovery
Continuing Calibration Verification 5 (5.00ppm Pb)	95.0% Recovery
Laboratory Control Standard	100% Recovery
Matrix Spike	97.6% Recovery
Duplicate Relative Percent Difference	2.43 RPD
Reporting Limit	20ug
Method Detection Limit	3.0ug

PREPARATION METHOD: ASTM E 1979-98M
ANALYSIS METHOD: EPA SW846 7420

ANALYST: Aubrey Simonds

Reviewed By Authorized Signatory:



Michael A. Mueller, MPH, Laboratory Director
Howard Varner, General Manager
Irma Faszewski, Quality Assurance Coordinator
David Xu, MS, Senior Chemist

Sample results denoted with a "less than" (<) sign contain less than 20ug total lead, based on a 40ml sample volume.

The condition of the samples analyzed was acceptable upon receipt per laboratory protocol unless otherwise noted on this report. Results represent the analysis of samples submitted by the client. Sample location, description, area, volume etc., was provided by the client. If the report does not contain the result for a field blank, it is due to the fact that the client did not include a field blank with their samples. EHS sample results do not reflect blank correction. This report shall not be reproduced, except in full, without the written consent of Environmental Hazards Services, L.L.C. California Certification #2319 NY ELAP #11714

LEGEND ug = microgram ppm = parts per million ug/ft² = micrograms per square foot
ml = milliliter Pb = lead ft² = square foot

wipepb15.AA220.doc/13MAR2007/REV3/MR/